



Perkin-Elmer HGA- 2100, based on the use of a 20 uL injection, continuous flow purge gas and non-pyrolytic graphite.

- 6.2 The use of background correction is recommended.
- 6.3 For every sample matrix analyzed, verification is necessary to determine that method of standard addition is not required (see part 5.2.1 of the Atomic Absorption Methods section of this manual).
- 6.4 If method of standard addition is required, follow the procedure given earlier in part 8.5 of the Atomic Absorption Methods section of this manual.
- 6.5 Data to entered into STORET must be reported as  $\mu\text{g/L}$ .

## 7.0 Precision and Accuracy

- 7.1 Precision and accuracy data are not available at this time.